



NC315 VGA Video Monitor Specification Sheet

IL752
Section E
Rev. 2 - 11/2009



NC315 VGA Video Monitor

ARCHITECTS' AND ENGINEERS' SPECIFICATION

The video monitor shall be TekTone® NC315 or approved equal. The video monitor shall be capable of tracking all Tek-CARE®NC300II system activities, including the display of all types and exact locations of calls. All calls shall be displayed in standard English format. The video monitor shall be standard VGA input.

The monitor shall have an anti-glare 15.1" LCD screen with a 15" display area. The monitor shall have a dot pitch of 0.3 mm with a resolution of 1024×768 (non-interlaced). The monitor shall provide controls for on/off with indicator, brightness, contrast, horizontal size, horizontal position, vertical size and vertical position.

The NC315 shall be ETL listed to the UL® 1069 standard.

FEATURES

The NC315 Video Monitor is used in conjunction with the NC304 and NC304LCD Nurses' Master Stations to provide complete control and communication functions, as offered by the Tek-CARE®NC300II Nurse Call System.

The video monitor visually follows all incoming calls and system activities. Up to 21 calls, identified by room and bed number, may be displayed simultaneously. Twelve different call priorities are indicated, and code and emergency calls are easily distinguished from routine calls. Nurse/aide-needed, nurse/aide-registered, and **STAT** staff emergency requests are also indicated. Calls waiting and overtime messages are provided as reminders that staff response is needed.

The 15" non-glare display area affords the operator clear visual indication of system activity. The video monitor is desk mounted.

- ETL listed to the UL® 1069 standard
- 15" LCD screen; 15" non-glare viewing area
- 0.3 mm dot pitch
- 1024×768 resolution (non-interlaced)
- Digital controls
- Low radiation

SPECIFICATIONS

Power:	110–220V, 50/60Hz (automatically detects power)	
Dimensions:	Height:	15" (381mm)
	Width:	14" (356mm)
	Depth:	8" (203mm)
Environment:	Temperature:	0°C–35°C
	Humidity:	30–80%
Connector:	15 Pin D-Sub	

REQUIRED COMPONENTS

NC304 or NC304LCD	Nurses' Master Station
NC351(A) series	Central Equipment

www.tektone.com

277 Industrial Park Road • Franklin, NC 28734 • tektone@tektone.net

Phone: (828) 524-9967 • Fax: (828) 524-9968 • Tech Support: Choose option 2 • Sales: Choose option 3

TekTone®'s quality system is registered by UL® to the ISO 9001 standard. (Reference #10001510.)